

**State of California**  
Department of Food and Agriculture  
Division of Measurement Standards

Certificate Number: 5652-11

Page 1 of 2

***California Type Evaluation Program***  
***Certificate of Approval***  
***for Weighing and Measuring Devices***

**For:**

Meter Indicating Volume  
Cryogenic Liquid Flow Meter, Turbine Meter  
Model: TMXXXXX-X-X-X-X  
Flow Rate: (see below)

**Submitted by:**

Turbines, Inc.  
P.O. Box 933  
Altus, OK 73522  
Tel: 864-882-4544  
Fax: 864-882-4457  
Contact: Kevin Clark  
Email: [Kevin.c@turbinesincorporated.com](mailto:Kevin.c@turbinesincorporated.com)  
Website: [www.turbinesincorporated.com](http://www.turbinesincorporated.com)

**Standard Features and Options**

- Lightweight hydraulically balanced rotor
- Standard end fittings
- Standard pick-up coil (temperature range -430 ° to +450 °F)
- Constructed of stainless steel

<b>Flow Meter Size</b>	<b>(GPM)</b>	<b>lbs/min</b>
TMC0075	7.5 to 38	55 to 300
TMC0100	7.5 to 60	55 to 500
TMC0150	8 to 130	60 to 1000
TMC0200	15 to 225	115 to 1700
TMC0300	40 to 380	300 to 3000

NOTE: All meters must be permanently marked to show the designated maximum and minimum discharge rates.

This device was evaluated under the California Type Evaluation Program (CTEP) and was found to comply with the applicable requirements of California Code of Regulations for "Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.



Effective Date: April 8, 2011

Kristin J. Macey, Director

**Turbines, Inc.**  
**Cryogenic Liquid Flow Meter, Turbine Meter**  
**Model: TMXXXXX-X-X-X-X**

**Application:** These meters are suitable for stationary and vehicle tank meter applications. This meter may be used with approved and compatible electronic indicating registers.

**Identification:** The required information is stamped on the metal body of the meter.

**Model Designation:**

Basic Model	Product Family	XXXX (Flow Meter Size)	X (End Fitting Type)	X (Bearing Type)	X (Rotor Type)	X (Pick-up Coils)
TM	Blank = Standard C = Cryogenics	0075 = 3/4" 0100 = 1" 0150 = 1 1/2" 0200 = 2" 0300 = 3"	Blank = 37° AN Flare NPT = Male National Pipe Thread F150 = 150# Flanges F300 = 300# Flanges F600 = 600# Flanges F900 = 900# Flanges F1500 = 1500# Flanges F2500 = 2500# Flanges W = Wafer Style VIC = Grooved Victaulic HP = Weco Detachable Nut GR = Grayloc Flanges	Blank = Ball Bearing Self-Lubricating T = Tungsten Carbide S = Special, Consult Factory	Blank = Nickel 200 PH = 17-4 PH DS = Duplex Stainless Steel	Blank = One Magnetic Coil 2Mag = Two Magnetic Coils 1Hub = One Coil, 1" MNPT Riser 2Hub = One Coil, 1" MNPT Riser

**Sealing:** The pick-up coil and temperature probe are secured individually by a wire security seal threaded through holes in the respective connector backshell and threaded coupling nut.

**Operation:** As product flows through a turbine flowmeter, rotor blades inside the flowmeter rotate through the magnetic field of the magnetic pick-up coil, thereby generating electrical pulses. The number and frequency of these pulses corresponds to the amount of product and flow rate respectively. The compensator corrects the value for temperature/density changes and sends the information to the register/ticket printer.

**Test Conditions:** A two inch and a three-quarter inch meters with a CDS1000 electronic indicator were installed on a liquid carbon dioxide trailer and a two inch meter installed on a liquid nitrogen trailer. Five tests at each of four flow rates were performed and a required throughput was approximately 1,069,518 gallons. The turbine meters were tested volumetrically and by mass. All initial and permanence test results were within the applicable tolerances of 1.5% and repeatability requirements.

Results of the evaluations indicate the devices comply with applicable requirements.

**Type Evaluation Criteria Used:** Title 4, California Code of Regulations, 2011 Edition

**Tested By:** A. Katalinic (NC), C. Nelson (CA)

**Example of Device:**



Model TMC0200